

Missouri Department of Transportation
Roberta Broeker, Interim Director

573.751.2551
Fax: 573.751.6555
1.888.ASK MODOT (275.6636)

ADDENDUM 001
Hamburg Wheel Tracking Device
RFB 5-150624DP

Offerors should acknowledge receipt of Addendum 001 by **signing** and **including it** with the original proposal. The due date for receipt of proposals is **unchanged** by this Addendum. Please see Attachment A, for further clarifications. All other terms and conditions remain unchanged and in full force.

QUESTIONS SUBMITTED BY BIDDERS

Please refer to Attachment A of this addendum for answers to the questions from bidder(s) regarding this solicitation.

Accordingly, this information **is believed to be of general interest to all potential Bidders.**

Name and Title of Signer (Print or type)	Name and Title of Department Authority Dee Pardue Sr. GS Technician
Contractor/Offeror Signature <hr/> (Signature of person authorized to sign)	Department of Transportation <i>Dee Pardue</i> <hr/> (Authorizing Signature)
Date Signed:	Date Signed: 6/17/2015



Our mission is to provide a world-class transportation experience that delights our customers and promotes a prosperous Missouri.

www.modot.org

QUESTIONS:

- 1. Q: Our unit offers a 1 year warranty and have never been asked for a 5 year warranty. Is this a requirement for acceptance of the bid? We are glad to extend it to 2 or 3 years, but I need to know how much weight this portion of the bid has on the final decision. Will our unit be disqualified if we do not offer a 5 year warranty?**

A: Yes, a 5-year warranty is an expected requirement of the bid. This was a decision made by the Construction & Materials group. It was made with the understanding it may increase the price of the product.

- 2. Q: The bid requests two calibrations. Is the first verification/calibration performed at installation and the second one performed a year later or is the installation verification not count toward the 2 calibration requested? Can you please clarify?**

A: Per the requirement of the bid, Section 5 – Calibration – “Vendor shall provide all necessary calibration/verification on all equipment and parts for the first two (2) years.” This should include all necessary calibrations needed to keep the unit within compliance for the first 2 years following installation.